## CERTIFICATE OF CALIBRATION

Calibrated for: Advance Devices, Inc. Certificate No: 16468 Serial Number: 21155 Software Version: R.1.04 Temperature: 23±2C Relative Humidity: 50+20% Model: ST5S Procedure Used: 4/6RES3.07

Certificate Date: 5/31/2015 9:10:23 PM

# This certified that above product was calibrated using applicable procedure.

As received condition: Factory tested

As shipped condition: At the completion of calibration this product meets published specification

Special Requirements: Re-certification of calibration will be performed upon request

Calibration Equipment Used: Smart Tweezers ST5S Calibration Module LVC139-CAL

(Certificate of calibration #136455 by Navair Technologies, Inc.)

#### **Accuracy Specification**

Measurement Range	Basic Measurement Accuracy
100 $\Omega$ to 10 k $\Omega$	Better than 0.2%
$0.1~\Omega$ to $9.9~M\Omega$	Better than 0.7%
10 nF to 100 μF	Better than 0.5%
100 pF to 1000 μF	Better than 2.0%
1 μH to 1 H	Better than 0.5%
1.0 μH to 999 mH	Better than 1.0%
	100 Ω to 10 kΩ 0.1 Ω to 9.9 MΩ 10 nF to 100 μF 100 pF to 1000 μF 1 μH to 1 H

### Maximum measurement ranges

Dissipation factor D: 0.001 to 1000

#### **Maximum resolution**

Resistance: 10 m $\Omega$  Capacitance: 0.1 pF Inductance: 0.1  $\mu$ H

2-wire measurements may introduce precision uncertainty up to 0.1%

Parameter	Measurement Range	Test frequency
Resistance	0.05 $\Omega$ to 9.9 $M\Omega$	1 kHz
Capacitance	0.5 pF to 999 pF 1000 pF to 1 μF > 1 μF	10 kHz 1 kHz 100 Hz
Inductance	0.5 μH to 999 μH 1 mH to 99 mH > 100mH	10 kHz 1 kHz 100 Hz

### **Typical offset:**

Resistance:  $25 \text{ m}\Omega$  Capacitance: 0.35 pF Inductance:  $0.1 \text{ }\mu\text{H}$ 

Offset value should be subtracted from measurement result for small values ( $R < 10\Omega$ , C < 100 pF, L < 10  $\mu$ H)

<sup>\*</sup> with 4-wire bench calibration at optimum test frequencies, ranges, DUT value, without offset.